

X-spine Systems, Inc.
Silex™ Sacroiliac Joint Fusion System

⚠ IMPORTANT NOTE:

The user acknowledges that he/she has read and agreed to the conditions in this insert, which are to be considered as contractual.

GENERAL INFORMATION

The Silex Sacroiliac Joint Fusion System consists of different diameter bone screws in various lengths and thread configurations to accommodate variations in patient anatomy. The Silex Sacroiliac Joint Fusion System is manufactured from Titanium alloy in accordance with ASTM F136 as well as an optional version where exterior surfaces are coated with medical-grade commercially pure titanium (CP Ti) per ASTM F1580. All implants are intended as single use only and should not be reused under any circumstances.

INDICATIONS FOR USE

The Silex Sacroiliac Joint Fusion System is intended for sacroiliac joint fusion for conditions including degenerative sacroiliitis and sacroiliac joint disruptions.

CONTRAINDICATIONS

Contraindications for the Silex Sacroiliac Joint Fusion System are similar to those of other systems of similar design, and include, but are not limited to:

1. Patients with probable intolerance to the materials used in the manufacture of this device.
2. Patients with infection, inflammation, fever, tumors, elevated white blood count, obesity, pregnancy, mental illness and other medical conditions which would prohibit beneficial surgical outcome.
3. Patients resistant to following post-operative restrictions on movement, especially in athletic and occupational activities.
4. Use with components from other systems.
5. Grossly distorted anatomy caused by congenital abnormalities.
6. Any other medical or surgical condition which would preclude the potential benefit of spinal implant surgery.
7. Rapid joint disease, bone absorption, osteopenia. Osteoporosis is a relative contraindication since this condition may limit the degree of obtainable correction, stabilization, and/or the amount of mechanical fixation.
8. Any case where the implant components selected for use would be too large or too small to achieve a successful result.
9. *Any patient having inadequate tissue coverage over the operative site or inadequate bone stock or quality.*

10. Any patient in which implant utilization would interfere with anatomical structures or expected physiological performance.
11. Any case not described in the indications for use.
12. Reuse or multiple uses.

WARNINGS AND PRECAUTIONS

As with any surgical system, the Silex Sacroiliac Joint Fusion system should be used by experienced surgeons with specific training in the use of the spinal system because this is a technically demanding procedure presenting a risk of serious injury to the patient.

Knowledge of surgical techniques, proper reduction, selection and placement of implants, and pre- and post-operative patient management are considerations essential to a successful surgical outcome. Appropriate selection, placement and fixation of the spinal system components are critical factors which affect implant service life. As in the case of all prosthetic implants, the durability of these components is affected by numerous biologic, biomechanics and other extrinsic factors, which limit their service life. Accordingly, strict adherence to the indications, contraindications, precautions, and warnings for this product is essential to potentially maximize service life. (Note: While proper implant selection can minimize risks, the size and shape of human bones present limitations on the size, shape, and strength of the implants).

Patients who smoke have been shown to have an increased incidence of pseudoarthrosis. Such patients should be advised of this fact and warned of the potential consequences. Patients with previous spinal surgery at the level to be treated may have different clinical outcomes compared to those without a previous surgery. Based on the fatigue testing results, the physician/surgeon should consider the level of implantation, patient weight, patient activity level, and other patient conditions, etc. which may have an impact on the performance of the system. For women of childbearing potential: If pregnancy occurs, the woman should review delivery options with her obstetrician. Vaginal delivery of a fetus may not be advisable following SI joint fusion.

If the patient is involved in an occupation or activity which applies inordinate stress upon the implant (e.g. substantial walking, running, lifting, or muscle strain) resultant forces can cause failure of the device. In some cases, progression of degenerative disease may be so advanced at the time of implantation that the expected useful life of the appliance may be substantially decreased. In such cases, orthopedic devices may be considered only as a delaying technique or to provide temporary relief. Patients should be instructed in detail about the limitations of the implants, including, but not limited to, the impact of excessive loading through patient weight or activity, and be taught to govern their activities accordingly. The patient should understand that a metallic implant is not as strong as normal, healthy bone and will bend, loosen or fracture if excessive demands are placed on it. An active, debilitated, or demented patient who cannot properly use weight supporting devices may be particularly at risk during postoperative rehabilitation.

Care must be taken to protect the components from being marred, nicked or notched as a result of contact with metal or abrasive objects. Alterations will produce defects in surface finish and internal stresses which may become the focal point for eventual breakage of the implant.

As with all orthopedic and neurosurgical implants, none of the Silex Sacroiliac Joint Fusion System components should ever be reused under any circumstances. Risks associated with reuse include infection, non-union (pseudarthrosis), serious patient injury or death.

Due to the presence of implants, interference with roentgenographic, CT and/or MR imaging may result. The Silex Sacroiliac Joint Fusion System has not been evaluated for safety and compatibility in the MR environment. The Silex Sacroiliac Joint Fusion System has not been tested for heating or migration in the MR environment. It must be noted that there are several different manufacturers and generations of MRI systems available, and X-spine cannot make any claims regarding the safety of X-spine implants and devices with any specific MR system.

Physician Note: The physician is the learned intermediary between the company and the patient. The indications, contraindications, warnings, and precautions given in this document must be conveyed to the patient. If requested, additional information, including surgical technique manuals, may be obtained through corporate sales representatives.

PREOPERATIVE MANAGEMENT

1. The surgeon should consider for surgery only those patients indicated for the use of this device.
2. The surgeon should not consider for surgery those patients contraindicated for the use of this device.
3. The surgeon should have a complete understanding of the device's indications, contraindications, and applications.
4. The surgeon should have a complete understanding of the function and limitations of each implant and instrument.
5. Device components should be received and accepted only in packages that have not been damaged or tampered with. Damaged implants and/or instruments should not be used. Components must be carefully handled and stored in a manner that prevents scratches, damage, and corrosion.
6. The type of implant to be used for the case should be determined prior to beginning the surgery.
7. All parts should be cleaned and sterilized before use.

INTRAOPERATIVE MANAGEMENT

1. Extreme caution should be used around the spinal cord and nerve roots. Damage to these structures will cause loss of neurological function.
2. Breakage, slippage, or misuse of instruments or implant components may cause injury to the patient or operative personnel.
3. Whenever possible or necessary, an imaging system should be utilized to facilitate surgery.

4. Caution should be taken in handling the implants; Damage to the implants may affect their performance.
5. Implants should not be reused under any circumstances.

INSTRUCTIONS FOR USE

For complete instructions regarding the proper use and application of all Silex implants and instruments, please refer to the Silex Surgical Technique Manual (available at no charge upon request).

POSTOPERATIVE MANAGEMENT

Postoperative management by the surgeon, including instruction and warning to and compliance by the patient, of the following is essential:

1. The patient should have a complete understanding of and compliance with the purpose and limitations of the implant devices.
2. Postoperative patients should be instructed to limit activity.
3. Rigid external orthosis/bracing should be utilized until fusion is confirmed clinically and radiographically.
4. During explantation, care should be taken to avoid damaging the implant and surrounding tissue as little as possible. The explanted device should be cleaned and disinfected using the instructions provided for cleaning/disinfection of instruments. Information on the procedure and patient should be retained to assist in any investigation.
5. Retrieved implants should be properly disposed of and are not to be reused under any circumstances.

POTENTIAL COMPLICATIONS AND ADVERSE SIDE EFFECTS

Potential complications and adverse effects for this system are similar to those of other spinal instrumentation systems, and include, but are not limited to:

1. Early or late loosening of any or all of the components.
2. Disassembly, bending, and/or breakage of any or all of the components.
3. Foreign body (allergic) reaction to implants.
4. Post-operative change in spinal curvature, loss of correction, height, and/or reduction.
5. Infection.
6. Dural tears, persistent CSF leakage, meningitis.
7. Loss of neurological function including paralysis (partial or complete), radiculopathy, and/or the development or continuation of pain, numbness, spasms, or sensory loss.
8. Cauda equina syndrome, neurological deficits, paraplegia, reflex deficits, irritation, and/or muscle loss.
9. Loss of bladder control or other types of urological system compromise.
10. Scar formation possibly causing neurological compromise or compression around nerves and/or pain.

11. Fracture, micro-fracture, resorption, damage, or penetration of any spinal bone.
12. Herniated nucleus pulposus, disc disruption or degeneration at, above, or below the level of surgery.
13. Non-union (pseudarthrosis), delayed union, mal-union.
14. Cessation of any potential growth of the operated portion of the spine.
15. Loss of or increase in spinal mobility or function.
16. Inability to perform the activities of daily living.
17. Death.

PACKAGING, LABELING, AND STORAGE


The implants are supplied clean and NON-STERILE. They must be sterilized (see below). The implants are delivered in packages. These must be intact at the time of receipt. All the legal information required for this type of implant is given on the label of each package. The implants may be delivered as a complete set: Implants and instruments are contained within specially designed trays or in boxes which can be sterilized directly. Use care in handling and storage of the implant components. Cutting, sharply bending, or scratching the surface can significantly reduce the strength and fatigue resistance of the implant system. This, in turn, could induce cracks and/or non-visible internal stresses that could lead to fracture of the implants. Implants and instruments in storage should be protected from corrosive environments such as salt, air, moisture, etc. Inspection and trial assembly are recommended prior to surgery to determine if instruments or implants have been damaged during the storage processes.

STERILIZATION

All Silex Sacroiliac Joint Fusion System implants and instruments are provided non-sterile and must be sterilized before use. All implants and instruments must be free of packaging material and bio-contaminants prior to sterilization. To achieve a sterility assurance level of not less than 10^{-6} , all non-sterile implants and instruments should be autoclave sterilized using the following validated cycle parameters:

Saturated steam method, pre-vacuum air removal, 270° F (132° C), 4-minute minimum exposure time, 30-minute minimum dry time, in a double-wrapped case configuration.

CLEANING OF INSTRUMENTS

 Caution: Use of sodium hydroxide (NaOH) is prohibited. Use of corrosive products and/or instruments including abrasive sponges and metal brushes should be avoided. Cleaning must be performed by personnel trained in the general procedures involving contaminant removal. Automated washer/disinfector systems are not recommended as the sole cleaning method for surgical instruments. An automated system may be used in addition to the following manual cleaning procedure.

1. Thoroughly clean all instruments prior to use and as soon as possible after use (within a maximum of 2 hours post-operation) with intensive rinsing under cool tap water (<40°C) to remove gross soil. Do not allow blood and debris to dry on the instruments. If cleaning must be delayed, place

instruments in a covered container with appropriate detergent (Enzol® Enzymatic Detergent or equivalent) to delay drying.

2. No instruments within this system require disassembly as part of the cleaning process.
3. The following table describes the required steps for thoroughly cleaning the system instruments:

Step	Agent	Minimum Time (mm:ss)
	Instructions	
1. Initial Clean	Enzol Enzymatic Detergent Solution (or equivalent)	10:00
	Add one (1) ounce (30 mL) of Enzol to one (1) gallon (3.8 L) of tap water. Soak instruments immediately after use and flush detergent through all channels until evidence of organic material is removed. Soak for a minimum of ten (10) minutes. Use a soft bristle brush (Spectrum™ M-16 or equivalent) to gently remove visible debris. Pay close attention to threads, crevices, lumens and hard to reach areas. If organic material is dried-on, extend soak time and use two (2) ounces (60 mL) of Enzol per one (1) gallon (3.8 L) of warm tap water.	
2. Rinse	Deionized water	3:00
	Thoroughly rinse each instrument with deionized water including all channels to remove detergent for a minimum of three (3) minutes.	
3. Inspection	Unaided eye	1:00
	Inspect each instrument for evidence of organic material. Particular attention should be taken to remove all debris from instruments with cannulations, holes, and features that may be shielded from brushing action. Subject instruments to ultrasonic cleaning if organic matter is present after the initial cleaning step.	
4. Ultrasonic Clean (if required)	Enzol Enzymatic Detergent Solution (or equivalent)	10:00
	Prepare a fresh solution by adding one (1) ounce (30 mL) of Enzol and one (1) gallon (3.8 L) of warm tap water to a sonication unit (Branson Bransonic® Ultrasonic Cleaner or equivalent). Fully immerse the instruments in the solution and sonicate for a minimum of ten (10) minutes.	
5. Ultrasonic Rinse	Deionized water	3:00
	Thoroughly rinse each instrument with deionized water including all holes and cannulations to remove detergent for a minimum of three (3) minutes.	
6.	Unaided eye	1:00

Inspection	Inspect each instrument for evidence of organic material. Repeat the ultrasonic clean and rinse steps if needed.
------------	--

4. Upon completion, visually inspect each instrument for contamination such as remaining soil and moisture or wetness. If soil remains, repeat the cleaning process. If wetness remains, use filtered pressurized air or lint-free wipes to dry.

INSPECTION

1. Carefully inspect each instrument to ensure all visible blood and soil has been removed.
2. Inspect instruments and instrument cases for damage. Check action of moving parts to ensure proper operation.
3. If damage or wear is noted that may compromise the proper function of the instrument or instrument case, do not use and contact customer service or your X-spine Systems representative for a replacement.
4. If corrosion is noted, do not use and contact customer service or your X-spine Systems representative for a replacement.

Manufacturer:




X-spine Systems, Inc.
 452 Alexandersville Rd.
 Miamisburg, OH 45342 USA
 Phone: (800) 903-0640
 Fax: (937) 847-8410


Authorized Representative:



EMERGO EUROPE
 Molenstraat 15
 2513 BH, The Hague
 The Netherlands
 Phone: +31.70.345.8570
 Fax: +31.70.346.7299

 **CAUTION:** Federal Law (USA) restricts these devices to use by or on the order of a physician.

CE 0086

Part Number
 X079-2003 D